
Solid State Drive Kit

This is the User's Guide for Solid State Drive Kit of E5081A ENA-X Series Network Analyzer.

Notices

© Keysight Technologies 2025

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Keysight Technologies, Inc. as governed by United States and international copyright laws.

Trademark Acknowledgments

Manual Part Number

E5081-90101

Edition

Edition 2, November 2025

Printed in Malaysia

Published by:

Keysight Technologies International Japan G.K.
1-3-3 Higashikawasaki-cho
Chuo-ku
Kobe-shi, Hyogo, Japan

Warranty

THE MATERIAL CONTAINED IN THIS DOCUMENT IS PROVIDED "AS IS," AND IS SUBJECT TO BEING CHANGED, WITHOUT NOTICE, IN FUTURE EDITIONS. FURTHER, TO THE MAXIMUM EXTENT PERMITTED BY APPLICABLE LAW, KEYSIGHT DISCLAIMS ALL WARRANTIES, EITHER EXPRESS OR IMPLIED WITH REGARD TO THIS MANUAL AND ANY INFORMATION CONTAINED HEREIN, INCLUDING BUT NOT LIMITED TO THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. KEYSIGHT SHALL NOT BE LIABLE FOR ERRORS OR FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH THE FURNISHING, USE, OR PERFORMANCE OF THIS DOCUMENT OR ANY INFORMATION CONTAINED HEREIN. SHOULD KEYSIGHT AND THE USER HAVE A SEPARATE WRITTEN AGREEMENT WITH WARRANTY TERMS COVERING THE MATERIAL IN THIS

DOCUMENT THAT CONFLICT WITH THESE TERMS, THE WARRANTY TERMS IN THE SEPARATE AGREEMENT WILL CONTROL.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

Declaration of Conformity

Declarations of Conformity for this product and for other Keysight products may be downloaded from the Web. Go to <http://www.keysight.com/go/conformity>. You can then search by product number to find the latest Declaration of Conformity.

U.S. Government Rights

The Software is "commercial computer software," as defined by Federal Acquisition Regulation ("FAR") 2.101. Pursuant to FAR 12.212 and 27.405-3 and Department of Defense FAR Supplement ("DFARS") 227.7202, the U.S. government acquires commercial computer software under the same terms by which the software is customarily provided to the public. Accordingly, Keysight provides the Software to U.S. government customers under its standard commercial license, which is embodied in its End User License Agreement (EULA), a copy of which can be found at <http://www.keysight.com/find/swela>. The license set forth in the EULA represents the exclusive authority by which the U.S. government may use, modify, distribute, or disclose the Software. The EULA and the license set forth therein, does not require or permit, among other things, that Keysight: (1) Furnish technical information related to commercial computer software or commercial computer software documentation that is not customarily provided to the public; or (2) Relinquish to, or otherwise provide, the government rights in excess of these rights customarily provided to the public to use, modify, reproduce, release, perform, display, or disclose

commercial computer software or commercial computer software documentation. No additional government requirements beyond those set forth in the EULA shall apply, except to the extent that those terms, rights, or licenses are explicitly required from all providers of commercial computer software pursuant to the FAR and the DFARS and are set forth specifically in writing elsewhere in the EULA. Keysight shall be under no obligation to update, revise or otherwise modify the Software. With respect to any technical data as defined by FAR 2.101, pursuant to FAR 12.211 and 27.404.2 and DFARS 227.7102, the U.S. government acquires no greater than Limited Rights as defined in FAR 27.401 or DFARS 227.7103-5 (c), as applicable in any technical data.

Safety Notices

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a CAUTION notice until the indicated conditions are fully understood and met.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a WARNING notice until the indicated conditions are fully understood and met.

Table of Contents

1. General Information	
Solid State Drive Spare Kit	5
Contents of the Upgrade Kit	5
2. Option Installation	
Introduction	7
Update Procedure for Solid State Drive	8
Installation Requirement	8
Backing Up Option Data	8
Removal Procedure	8
Replacement Procedure	9
Restoring Option Data	10
Writing System Correction Data of Backup Memory to Removable Drive	13
Calibration of the Touch Screen	13

1 General Information

This chapter describes the Solid State Drive Kit, contents of the kit, and required tools to install the SSD spare kit.

Solid State Drive Spare Kit

The purpose of this kit is to install the SSD spare kit.

Contents of the Upgrade Kit

The kit includes the parts listed in **Table 1-1**.

Table 1-1 **Contents of the Upgrade Kit**

Description	Qty
Upgrade Manual (This manual)	1
Solid State Drive (SSD) Spare Kit	1

General Information
Contents of the Upgrade Kit

2 Option Installation

This chapter describes the procedure to install the SSD spare kit.

Introduction

This chapter describes the procedure to install the SSD spare kit. The procedure consists of the following items.

- **Installation Requirement**
- **Backing Up Option Data**
- **Removal Procedure**
- **Replacement Procedure**
- **Restoring Option Data**
- **Writing System Correction Data of Backup Memory to Removable Drive**
- **Calibration of the Touch Screen**

Update Procedure for Solid State Drive

Installation Requirement

- Optional
 - Philips (Flat) screwdriver

Backing Up Option Data

Step 1. Copy “C:\Program Files\Agilent\licensing\” folder from the original SSD to USB pen drive.

NOTE

This is an optional but highly recommended step.

It enables users to have a backup copy of the licenses. Even if you lost the file(s), you can activate the option again if you have the license number.

You can re-create the license number at

<https://www.keysight.com/find/softwaremanager> with your software entitlement certificate.

NOTE

If you borrow floating licenses, you need to return them before removing SSD. Otherwise, the available count of the licenses are still consumed at a brand-new SSD.

Removal Procedure

Step 1. Turn off the power.

Step 2. With the use of Philips screw driver or user hand, turn the two screws anti-clockwise until the screws are fully released.



Step 3. Pull the housing to remove the solid-state drive.



Replacement Procedure

Step 1. Install the SSD assembly into the instrument. Ensure the SSD assembly is installed without gap.


Step 2. Hand-tighten (push and turn clock-wise direction) or fasten the two screws with Philips screw drivers.



Restoring Option Data

Step 1. Turn on the analyzer.

Step 2. The EULA (Software End User License Agreement) will be prompted. Press **[Agree]**.




Step 3. System will proceed with accounts creation followed with and multiple reboots. Wait until the process is completed.

Step 4. No action is required from user in restoring the option data. The options will be automatically restored from backup memory in Network Analyzer to the new removable SSD in the background upon firmware start.

Step 5. Proceed to launch the Firmware.


The firmware will check for the FPGA version. If the following message appears, click "Yes" to proceed with the FPGA update. Refer to the on screen instruction followed with the full system restart.



Option Installation

Update Procedure for Solid State Drive


Step 6. Verify the product model number, serial number, and options are correctly listed in <System> -> Help -> About NA.




NOTE

In case the required options aren't listed,

1. Open Keysight License Manager in Network Analyzer. Select Install License File.



2. Connect USB pen drive to network analyzer. Navigate to the folder the backup licenses are saved earlier.
3. To install the license file(s), select the license file(s) and click Open.



NOTE

If you return floating licenses at the step **“Backing Up Option Data”**, you need to borrow them again.

Writing System Correction Data of Backup Memory to Removable Drive

The system correction data (Syscalls) are stored in the Network Analyzer backup memory. It is not affected by the SSD replacement.

No action is needed.

Calibration of the Touch Screen

For Windows 10 based E5081A


NOTE

The step is optional.


The default Windows touch screen calibration setting will be accurate. Only perform if user would like to customize the touch screen to their requirement.

Refer to the procedure described below to calibrate the touch-screen.

Step 1. In Windows operating system, navigate to **Control Panel -> Hardware and Sound -> Tablet PC Settings**. Click **“Calibrate”**.



Step 2. Windows UAC (User Account control) will be prompted. Press **“Yes”**.



Step 3. Touch the + crosshair according to the sequence on screen.



For Windows 11 based E5081A

NOTE

The step is optional.

The default Windows touch screen calibration setting will be accurate.


Touch screen calibration is no longer required in Windows 11 since the touch screen and display panel is pre-calibrated.

Refer to the following procedure to verify the touch screen functionality.


Step 1. In Windows operating system, navigate to **Control Panel -> Hardware and Sound -> Tablet PC Settings**. Under **Display -> Configure**, click **“Setup”**.

NOTE


In newer releases of Windows 11, the Tablet PC Settings function might have been removed. Skip steps 1 and 2 if it is not applicable.




Step 2. The following screen would be prompted. Tap this screen with single finger to identify the touch screen functionality. Upon successful recognition of touch/tap input, the tool will automatically exit.



Step 3. If user would like to customize the touch screen to their preference, click **Start**. Enter “tabcal”. Launch the utility.



Step 4. Windows UAC would be prompted. Press “Yes”.



Step 5. Touch the + crosshair according to the sequence on screen.



Option Installation
Update Procedure for Solid State Drive

This information is subject to change without notice.

© Keysight Technologies 2025

Edition 2, November 2025



E5081-90101

www.keysight.com